

## Scope of Accreditation For Qualitex Inc.

4185 East Park 30 Drive  
Columbia City, IN 46725  
Rick Williams  
260-244-7839

In recognition of a successful assessment to ISO/IEC 17025:2005, accreditation is granted to **Qualitex Inc.** to perform **Calibration and Dimensional Inspection** as shown in the following tables:

Accreditation Granted Through: **December 30, 2012**

### Calibration

#### Length - Dimensional Metrology – Hand Tools and Precision Gages 1D

Calibration Parameter/Equipment	Range	Calibration and Measurement Capability(+/-) <sup>1,2</sup>	Remarks
Indicators Analog Dial and Digital	0 in to 2 in	(400 + 69L) μin	Gage Blocks

### Dimensional Inspection

#### Length - Dimensional Inspection – Dimensional Measurement 3D

Inspection Parameter	Range	Calibration and Measurement Capability (+/-) <sup>1,3</sup>	Remarks
Dimensional Measurement 3D Steel	X = 0 mm to 900 mm Y = 0 mm to 1500 mm Z = 0 mm to 800 mm	(9.6 + 0.006L) μm	Coordinate Measuring Machine utilized as Reference Standard for Dimensional Inspection
Dimensional Measurement 3D Aluminum	X = 0 mm to 900 mm Y = 0 mm to 1500 mm Z = 0 mm to 800 mm	(9.6 + 0.033L) μm	

Note(s):

- 1) Calibration and Measurement Capability represent expanded uncertainties at approximately the 95% confidence level using a coverage factor of k=2.
- 2) L = Length in inches
- 3) L = Length in millimeters

Approved by: \_\_\_\_\_

R. Douglas Leonard  
Chief Technical Officer

Date: December 30, 2009

Issued: 12/30/09